



# QA Expectations for M4020 Met

<b>Initial Calibration, 3 point</b>	<b>Every batch of 20</b>	<b><math>R^2 \geq 0.95</math></b>	<b>MET</b>
<b>Reference Material; Known Standard (NRCC HS-1)</b>	<b>Once</b>	<b>% Recovery = 70-130%</b>	<b>MET</b>
<b>Continuing Calibration Verification</b>	<b>Aroclor 1254 std. every batch of 20</b>	<b><math>R^2 \leq 0.95</math></b>	<b>MET</b>
<b>Method Blank (Methanol Extractant)</b>	<b>Every batch of 20</b>	<b><math>\leq RL</math></b>	<b>MET</b>
<b>Lab Duplicates</b>	<b>Every batch of 20</b>	<b>% RPD <math>\leq 30\%</math></b>	<b>MET</b>